


















Operator netty

				< Range >		Flags and Alarms			
 <p>RBC</p>	RBC	4.63	10 <sup>6</sup> /mm <sup>3</sup>	4.50	6.50	Analyzer Alarms			
	HGB	14.3	g/dL	13.0	17.0	CO			
	HCT	41.4	%	40.0	54.0	LMNE-			
	MCV	89	µm <sup>3</sup>	80	100	Suspected Pathology			
	MCH	30.8	pg	27.0	32.0	WBC Interpretation Not Possible			
	MCHC	34.5	g/dL	32.0	36.0	Remarks			
	RDWcv	11.8	%	11.0	16.0	RBC of the Run 19/07/2017 9:45:17			
	RDWsd	38	µm <sup>3</sup>	39	52	WBC of the Run 19/07/2017 9:45:17			
	PLT	297	10 <sup>3</sup> /mm <sup>3</sup>	150	500	PLT of the Run 19/07/2017 9:45:17			
	MPV	8.1	µm <sup>3</sup>	6.0	11.0	DIFF of the Run 19/07/2017 9:45:17			
PCT	0.239	%	0.150	0.500					
PDW	14.0	%	11.0	18.0					
 <p>PLT</p>	WBC	13.9	IH 10 <sup>3</sup> /mm <sup>3</sup>	4.0	10.0	 <p>LMNE</p>			
	NEU	87.3	% *h	12.09	% *H			50.0	70.0
	LYM	8.0	% *l	1.11	% *			30.0	45.0
	MON	2.9	% *	0.40	% *			2.0	8.0
	EOS	1.3	% *	0.18	% *			1.0	6.0
	BAS	0.5	% *	0.07	% *			0.0	2.0
	ALY	1.2	% *	0.16	% *			0.0	2.5
	LIC	0.6	% *	0.08	% *			0.0	3.0

Operator	Technician			< Range >		Flags and Alarms			
 <p>RBC</p>	RBC	5.13	10 <sup>9</sup> /mm <sup>3</sup>	4.50	6.50	<b>Morphology Flags</b> ALY <b>Suspected Pathology</b> Leukopenia Lymphopenia Neutropenia Atypical Lymphocyte Thrombocytopenia  <b>Remarks</b> RBC of the Run 06/07/2017 09:15:42 WBC of the Run 06/07/2017 09:15:42 PLT of the Run 06/07/2017 09:15:42 DIFF of the Run 06/07/2017 09:15:42			
	HGB	15.7	g/dL	13.0	17.0				
	HCT	44.6	%	40.0	54.0				
	MCV	87	µm <sup>3</sup>	80	100				
	MCH	30.6	pg	27.0	32.0				
	MCHC	35.2	g/dL	32.0	36.0				
	RDWcv	11.8	%	11.0	16.0				
	RDWsd	37	L µm <sup>3</sup>	39	52				
	PLT	8	L 10 <sup>3</sup> /mm <sup>3</sup>	150	500				
	MPV	7.7	µm <sup>3</sup>	6.0	11.0				
 <p>PLT</p>	PCT	0.006	I %	0.150	0.500				
	PDW	17.0	%	11.0	18.0				
	 <p>BASO</p>	WBC	2.4	L 10 <sup>3</sup> /mm <sup>3</sup>	4.0	10.0			
			%	#	< %	Range	#	>	
		NEU	57.0	1.35	L	50.0	70.0	2.00	7.50
		LYM	24.1	0.57	L	30.0	45.0	1.00	4.00
		MON	14.8	0.35		2.0	8.0	0.20	1.00
		EOS	0.8	0.02		1.0	6.0	0.00	0.50
		BAS	3.3	0.08		0.0	2.0	0.00	0.20
		ALY	2.7	0.07	H	0.0	2.5	0.00	0.25
LIC	2.1	0.05		0.0	3.0	0.00	0.30		
				 <p>LMNE</p>					

Operator	Technician			< Range >		Flags and Alarms					
 <p>RBC</p>	RBC	5.35	10 <sup>6</sup> /mm <sup>3</sup>	4.50	6.50	<b>Analyzer Alarms</b> CO NO LMNE- <b>Suspected Pathology</b> WBC Interpretation Not Possible Platelet Aggregates <b>Remarks</b> RBC of the Run 06/07/2017 09:25:15 WBC of the Run 06/07/2017 09:25:15 PLT of the Run 06/07/2017 09:25:15 DIFF of the Run 06/07/2017 09:25:15					
	HGB	15.6	g/dL	13.0	17.0						
	HCT	43.8	%	40.0	54.0						
	MCV	82	µm <sup>3</sup>	80	100						
	MCH	29.1	pg	27.0	32.0						
	MCHC	35.5	g/dL	32.0	36.0						
	RDWcv	11.5	%	11.0	16.0						
	RDWsd	34	L µm <sup>3</sup>	39	52						
	PLT	136	II 10 <sup>3</sup> /mm <sup>3</sup>	150	500						
	MPV	7.9	µm <sup>3</sup>	6.0	11.0						
	PCT	0.108	I %	0.150	0.500						
	PDW	13.0	%	11.0	18.0						
 <p>PLT</p>	WBC	3.3	II 10 <sup>3</sup> /mm <sup>3</sup>	4.0	10.0	 <p>LMNE</p>					
		%	#	< %	Range			#	>		
	NEU	73.5	*h	2.43	*			50.0	70.0	2.00	7.50
	LYM	18.2	*l	0.60	*L			30.0	45.0	1.00	4.00
	MON	6.4	*	0.21	*			2.0	8.0	0.20	1.00
	EOS	1.5	*	0.05	*			1.0	6.0	0.00	0.50
	BAS	0.4	*	0.01	*			0.0	2.0	0.00	0.20
	ALY	1.7	*	0.06	*			0.0	2.5	0.00	0.25
LIC	3.0	*	0.10	*	0.0	3.0	0.00	0.30			

Operator		Technician		< Range >		Flags and Alarms																																																																			
 <p>RBC</p>	RBC	4.79	10 <sup>6</sup> /mm <sup>3</sup>	3.80	5.80	<b>Suspected Pathology</b> Neutrophilia  <b>Remarks</b> RBC of the Run 07/07/2017 9:25:55 WBC of the Run 07/07/2017 9:25:55 PLT of the Run 07/07/2017 9:25:55 DIFF of the Run 07/07/2017 9:25:55																																																																			
	HGB	13.4	g/dL	11.5	16.0																																																																				
	HCT	39.2	%	37.0	47.0																																																																				
	MCV	82	μm <sup>3</sup>	80	100																																																																				
	MCH	28.1	pg	27.0	32.0																																																																				
	MCHC	34.3	g/dL	32.0	36.0																																																																				
	RDWcv	12.2	%	11.0	16.0																																																																				
	RDWsd	36	L μm <sup>3</sup>	39	52																																																																				
	PLT	292	10 <sup>3</sup> /mm <sup>3</sup>	150	500																																																																				
	MPV	8.8	μm <sup>3</sup>	6.0	11.0																																																																				
 <p>PLT</p>	PCT	0.258	%	0.150	0.500	 <p>LMNE</p>																																																																			
	PDW	15.3	%	11.0	18.0																																																																				
	WBC	10.7	h 10 <sup>3</sup> /mm <sup>3</sup>	4.0	10.0			<table border="1"> <thead> <tr> <th></th> <th>%</th> <th>#</th> <th>&lt;</th> <th>%</th> <th>Range</th> <th>#</th> <th>&gt;</th> </tr> </thead> <tbody> <tr> <td>NEU</td> <td>78.9</td> <td>h 8.45</td> <td>H</td> <td>50.0</td> <td>70.0</td> <td>2.00</td> <td>7.50</td> </tr> <tr> <td>LYM</td> <td>13.0</td> <td>l 1.39</td> <td></td> <td>30.0</td> <td>45.0</td> <td>1.00</td> <td>4.00</td> </tr> <tr> <td>MON</td> <td>6.5</td> <td>0.70</td> <td></td> <td>2.0</td> <td>8.0</td> <td>0.20</td> <td>1.00</td> </tr> <tr> <td>EOS</td> <td>0.9</td> <td>l 0.10</td> <td></td> <td>1.0</td> <td>6.0</td> <td>0.00</td> <td>0.50</td> </tr> <tr> <td>BAS</td> <td>0.7</td> <td>0.07</td> <td></td> <td>0.0</td> <td>2.0</td> <td>0.00</td> <td>0.20</td> </tr> <tr> <td>ALY</td> <td>1.0</td> <td>0.11</td> <td></td> <td>0.0</td> <td>2.5</td> <td>0.00</td> <td>0.25</td> </tr> <tr> <td>LIC</td> <td>0.6</td> <td>0.07</td> <td></td> <td>0.0</td> <td>3.0</td> <td>0.00</td> <td>0.30</td> </tr> </tbody> </table>			%	#	<	%	Range	#	>	NEU	78.9	h 8.45	H	50.0	70.0	2.00	7.50	LYM	13.0	l 1.39		30.0	45.0	1.00	4.00	MON	6.5	0.70		2.0	8.0	0.20	1.00	EOS	0.9	l 0.10		1.0	6.0	0.00	0.50	BAS	0.7	0.07		0.0	2.0	0.00	0.20	ALY	1.0	0.11		0.0	2.5	0.00	0.25	LIC	0.6	0.07		0.0	3.0	0.00	0.30
		%	#	<	%					Range	#	>																																																													
	NEU	78.9	h 8.45	H	50.0					70.0	2.00	7.50																																																													
	LYM	13.0	l 1.39		30.0					45.0	1.00	4.00																																																													
	MON	6.5	0.70		2.0					8.0	0.20	1.00																																																													
	EOS	0.9	l 0.10		1.0					6.0	0.00	0.50																																																													
	BAS	0.7	0.07		0.0					2.0	0.00	0.20																																																													
	ALY	1.0	0.11		0.0					2.5	0.00	0.25																																																													
LIC	0.6	0.07		0.0	3.0	0.00	0.30																																																																		
NEU	78.9	h	8.45	H	50.0	70.0	2.00			7.50																																																															
LYM	13.0	l	1.39		30.0	45.0	1.00	4.00																																																																	
MON	6.5		0.70		2.0	8.0	0.20	1.00																																																																	
EOS	0.9	l	0.10		1.0	6.0	0.00	0.50																																																																	
BAS	0.7		0.07		0.0	2.0	0.00	0.20																																																																	
ALY	1.0		0.11		0.0	2.5	0.00	0.25																																																																	
LIC	0.6		0.07		0.0	3.0	0.00	0.30																																																																	

Operator		Technician		< Range >		Flags and Alarms			
	RBC	4.73	$10^6/\text{mm}^3$	3.80	5.80	<b>Remarks</b> RBC of the Run 11/07/2017 09:11:39 WBC of the Run 11/07/2017 09:11:39 PLT of the Run 11/07/2017 09:11:39 DIFF of the Run 11/07/2017 09:11:39			
	HGB	14.2	g/dL	11.5	16.0				
	HCT	41.0	%	37.0	47.0				
	MCV	87	$\mu\text{m}^3$	80	100				
	MCH	30.0	pg	27.0	32.0				
	MCHC	34.6	g/dL	32.0	36.0				
	RDWcv	11.0	I %	11.0	16.0				
	RDWsd	34	L $\mu\text{m}^3$	39	52				
	PLT	505	$h \cdot 10^3/\text{mm}^3$	150	500				
	MPV	7.0	$\mu\text{m}^3$	6.0	11.0				
	PCT	0.355	%	0.150	0.500				
	PDW	10.0	I %	11.0	18.0				
	WBC	9.8	$10^3/\text{mm}^3$	4.0	10.0	<b>LMNE</b> 			
		%	#	< %	Range			# >	
	NEU	53.0	5.17	50.0	70.0			2.00	7.50
	LYM	38.5	3.76	30.0	45.0			1.00	4.00
	MON	5.6	0.55	2.0	8.0			0.20	1.00
	EOS	2.0	0.20	1.0	6.0			0.00	0.50
	BAS	0.9	0.09	0.0	2.0			0.00	0.20
	ALY	2.1	0.20	0.0	2.5			0.00	0.25
LIC	0.4	0.04	0.0	3.0	0.00	0.30			